



## CERTIFICATE OF COMPLIANCE

**Certificate No.:**

K136150.01

**Date of Issue:**

4 February 2014

Applicant Name:

MediTouch Ltd.

Applicant Address

4 Hamelacha St., Poleg Industrial Area, Netanya, 4250574, Israel

Product

Balance Rehabilitation Platform

Model / Type Ref.

BalanceTutor™ BT100

Rating

230 VAC

Tested to

IEC 60601-1-2: 2007 (See Notes on following page).

Test Results are detailed in the E136150.01  
Test Report No.

Name: I. Raz

Signature: \_\_\_\_\_

**Certification Dept.**

This is to certify that the product specified herein has been tested & the test results were found to be compliant with the requirements noted above.



## CERTIFICATE OF COMPLIANCE

**Certificate No.:**

K136150.01

**Date of Issue:**

4 February 2014

Applicant Name:

MediTouch Ltd.

Notes:

1. Testing was performed at the customer premises.
2. The following IEC 60601-1-2 tests were performed according to the customer's request:

CISPR 11	Conducted and radiated emissions
IEC 61000-4-2	ESD
IEC 61000-4-4	EFT/B
IEC 61000-4-6	Conducted disturbances
IEC 61000-4-11	Voltage dips