

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Platform for Static and Dynamic Balance Rehabilitation

Name and address of the applicant

MediTouch Ltd.
45 Hamelacha St., Netanya
Israel

Name and address of the manufacturer

MediTouch Ltd.
45 Hamelacha St., Netanya
Israel

Name and address of the factory

Note: When more than one factory, please report on page 2

☐ Additional information on page 2

MediTouch Ltd.
45 Hamelacha St., Netanya
Israel

Ratings and principal characteristics

220-240Vac, 50/60Hz, 10A

Trademark (if any)



MediTouch

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

BalanceTutor: Model BT100

Additional information (if necessary may also be reported on page 2)

☐ Additional information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005/AMD1:2012

National differences:

CA, US

As shown in the Test Report Ref. No. which forms part of this Certificate

CB190700.01

This CB Test Certificate is issued by the National Certification Body

I.T.L. (PRODUCT TESTING) Ltd.
1 Bat Sheva St., Lod 7120101, Israel

Date: 2018-10-17

Signature: Steil Loznen

page 1 of 1